
Midmark M11 - v1.0.0
Total Cycles: 3844

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2023 03:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.1 F 0.0
2:00 187.5 F 0.7
4:00 198.7 F 4.9
6:00 217.3 F 10.0
8:00 224.6 F 6.5
10:00 237.3 F 10.8
12:00 235.1 F 8.3
14:00 241.0 F 10.2
16:00 248.8 F 14.2
18:00 261.4 F 21.3
20:00 272.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.8
1:00 271.8 F 27.8
2:00 272.3 F 28.2
3:00 272.7 F 28.5
4:00 271.7 F 28.1

Min 271.5 F 27.7
Max 272.7 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 222.3 F 0.0
10:00 228.8 F 0.0
15:00 228.8 F 0.0
20:00 224.9 F 0.0
25:00 223.5 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 20:19
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

09/29/2023 04:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____