
Midmark M11 - v1.0.0
Total Cycles: 2746

Name: RITTER 1
Sterilizer ID: V2178198
01/06/2023 09:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.1 F 0.0
2:00 79.6 F 0.0
4:00 88.6 F 0.0
6:00 115.9 F 0.0
8:00 147.6 F 0.0
10:00 176.7 F 2.3
12:00 195.6 F 5.4
14:00 207.5 F 8.7
16:00 216.6 F 12.3
18:00 224.6 F 6.2
20:00 234.4 F 10.5
22:00 235.2 F 8.4
24:00 243.4 F 12.6
26:00 240.7 F 10.7
28:00 253.3 F 16.9
30:00 263.9 F 23.2
32:00 271.7 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.9
2:00 271.4 F 28.2
3:00 272.6 F 29.0
4:00 271.7 F 28.2

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 206.9 F 0.0
10:00 215.4 F 0.0
15:00 221.5 F 0.0
20:00 219.3 F 0.0
25:00 222.0 F 0.0
30:00 225.3 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 32:06
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:10:13
CYCLE COMPLETE

01/06/2023 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____