
Midmark M11 - v1.0.0
Total Cycles: 5141

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2024 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 73.1 F 0.0
4:00 110.6 F 0.0
6:00 127.0 F 0.0
8:00 155.8 F 0.3
10:00 182.8 F 0.4
12:00 200.3 F 1.1
14:00 217.8 F 4.5
16:00 225.4 F 7.7
18:00 235.4 F 11.4
20:00 231.5 F 7.4
22:00 244.6 F 13.2
24:00 242.6 F 11.6
26:00 240.2 F 10.4
28:00 254.4 F 17.4
30:00 266.0 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.9
2:00 272.4 F 28.5
3:00 271.5 F 28.0
4:00 272.6 F 28.9

Min 271.5 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 209.1 F 0.0
10:00 215.6 F 0.0
15:00 228.9 F 0.0
20:00 221.8 F 0.0
25:00 230.8 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:41
HEATING: 31:44
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:50
CYCLE COMPLETE

11/29/2024 10:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____