

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5051

Name: RITTER 1  
Sterilizer ID: V2178198  
11/04/2024 01:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	173.0 F	0.0
2:00	176.6 F	2.4
4:00	196.0 F	7.2
6:00	214.1 F	12.1
8:00	223.4 F	6.9
10:00	235.5 F	12.3
12:00	237.3 F	9.9
14:00	236.6 F	8.6
16:00	250.3 F	15.3
18:00	262.3 F	22.1
20:00	272.8 F	29.0

C010: POWER UP MODE  
SYSTEM PWR LOSS

\*\* ITEMS NOT STERILE \*\*

11/04/2024 01:51 PM