

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 003447  
8:42 14/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:07  
PRESSURIZING 2:10  
STERILIZING 3:06  
136.0 C 220kPa 3:06  
135.8 C 219kPa 3:36  
136.4 C 222kPa 4:06  
136.4 C 223kPa 4:36  
136.3 C 222kPa 5:06  
136.4 C 223kPa 5:36  
136.4 C 222kPa 6:06  
136.3 C 222kPa 6:36  
136.3 C 222kPa 6:36  
VENTING 6:36  
AIR DRYING 7:25  
CYCLE COMPLETE 34:13

-----  
Digital Signature #  
83A41EBD2C2CA5C5  
-----