
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000992
9:42 14/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:08
STERILIZING 3:09
136.4 C 220kPa 3:09
136.0 C 231kPa 3:39
136.3 C 230kPa 4:09
136.4 C 233kPa 4:39
136.4 C 233kPa 5:09
136.3 C 232kPa 5:39
136.4 C 233kPa 6:09
136.4 C 233kPa 6:39
136.4 C 233kPa 6:39
VENTING 6:40
AIR DRYING 7:33
CYCLE COMPLETE 9:16

Digital Signature #
CD20D7CE8F7716E8
