

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4730

Name: RITTER 1  
Sterilizer ID: V2178198  
07/15/2024 02:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.5 F 0.0  
2:00 152.9 F 2.9  
4:00 171.2 F 6.4  
6:00 193.3 F 10.5  
8:00 203.9 F 4.9  
10:00 220.8 F 7.5  
12:00 231.7 F 12.3  
14:00 233.4 F 8.9  
16:00 239.2 F 10.5  
18:00 245.2 F 13.1  
20:00 256.7 F 19.4  
22:00 267.9 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.8 F 29.6  
2:00 272.1 F 29.0  
3:00 272.0 F 29.3  
4:00 272.8 F 29.6

Min 271.5 F 28.8  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 212.7 F 0.0  
10:00 213.5 F 0.0  
15:00 216.0 F 0.0  
20:00 225.3 F 0.0  
25:00 236.9 F 0.0  
30:00 232.2 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 23:28  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:02:23  
CYCLE COMPLETE

07/15/2024 03:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_