
Midmark M11 - v1.0.0
Total Cycles: 3778

Name: RITTER 1
Sterilizer ID: V2178198
09/15/2023 07:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 71.4 F 0.0
4:00 83.3 F 0.0
6:00 115.9 F 0.0
8:00 145.5 F 1.2
10:00 173.1 F 2.8
12:00 186.6 F 5.5
14:00 207.0 F 8.7
16:00 220.0 F 12.1
18:00 222.9 F 5.1
20:00 231.7 F 9.0
22:00 241.0 F 13.2
24:00 238.4 F 9.9
26:00 238.4 F 9.0
28:00 247.2 F 13.6
30:00 258.5 F 19.7
32:00 268.9 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.5 F 28.6
2:00 271.3 F 28.0
3:00 272.5 F 28.7
4:00 271.5 F 28.0

Min 271.3 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 168.5 F 0.0
5:00 244.7 F 0.0
10:00 205.2 F 0.0
15:00 183.3 F 0.0
20:00 170.4 F 0.0
25:00 169.8 F 0.0
30:00 169.2 F 0.0

Drying Complete

FILLING: 0:45
HEATING: 33:10
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:11:16
CYCLE COMPLETE

09/15/2023 09:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____