
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005464
12:36 02/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:43
PRESSURIZING 0:50
STERILIZING 1:53
136.4 C 223kPa 1:53
136.2 C 222kPa 2:23
136.4 C 224kPa 2:53
136.3 C 224kPa 3:23
136.4 C 223kPa 3:53
136.4 C 221kPa 4:23
136.4 C 223kPa 4:53
136.4 C 224kPa 5:23
136.4 C 224kPa 5:23
VENTING 5:23
AIR DRYING 6:17
CYCLE COMPLETE 6:22

Digital Signature #
C9DCA236B556B6E5
