

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3431

Name: RITTER 1  
Sterilizer ID: V2178198  
06/22/2023 01:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.3 F 0.0  
2:00 137.7 F 0.0  
4:00 163.8 F 1.5  
6:00 186.7 F 5.1  
8:00 205.9 F 9.1  
10:00 221.2 F 10.2  
12:00 228.6 F 8.6  
14:00 239.9 F 13.5  
16:00 239.5 F 10.8  
18:00 236.7 F 8.9  
20:00 251.3 F 15.8  
22:00 262.8 F 22.4  
24:00 272.6 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.6 F 28.9  
2:00 271.7 F 28.2  
3:00 272.0 F 28.7  
4:00 272.4 F 28.8

Min 271.3 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
5:00 219.7 F 0.0  
10:00 228.8 F 0.0  
15:00 229.5 F 0.0  
20:00 228.2 F 0.0  
25:00 230.7 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 24:19  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:02:33  
CYCLE COMPLETE

06/22/2023 02:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_