
Midmark M11 - v1.0.0
Total Cycles: 2030

Name: RITTER 1
Sterilizer ID: V2178198
06/16/2022 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.9 F 0.0
2:00 177.3 F 2.9
4:00 190.4 F 7.1
6:00 213.9 F 12.4
8:00 227.7 F 7.5
10:00 240.9 F 12.8
12:00 238.2 F 10.2
14:00 235.2 F 8.0
16:00 249.7 F 15.2
18:00 261.7 F 21.9
20:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.7 F 29.2
2:00 271.6 F 28.3
3:00 271.9 F 29.0
4:00 272.4 F 28.9

Min 271.1 F 28.2
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 217.8 F 0.0
10:00 223.3 F 0.0
15:00 215.2 F 0.0
20:00 223.4 F 0.0
25:00 230.4 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

06/16/2022 12:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____