
Midmark M11 - v1.0.0
Total Cycles: 2618

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2022 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.4 F 0.0
2:00 196.5 F 3.5
4:00 211.8 F 8.0
6:00 228.8 F 13.1
8:00 228.7 F 7.9
10:00 238.7 F 13.5
12:00 240.1 F 10.9
14:00 237.7 F 9.3
16:00 252.2 F 16.3
18:00 264.3 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.1 F 28.6
2:00 271.4 F 28.4
3:00 272.1 F 28.6
4:00 271.2 F 28.2

Min 271.1 F 28.0
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 228.3 F 0.0
10:00 228.3 F 0.0
15:00 230.3 F 0.0
20:00 231.6 F 0.0
25:00 232.0 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

11/29/2022 11:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____