

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5019

Name: RITTER 1  
Sterilizer ID: V2178198  
10/24/2024 08:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.5 F 0.0  
2:00 79.7 F 0.0  
4:00 77.5 F 0.0  
6:00 128.4 F 0.1  
8:00 160.7 F 0.4  
10:00 175.3 F 0.8  
12:00 195.3 F 3.8  
14:00 218.3 F 7.1  
16:00 229.6 F 10.5  
18:00 231.9 F 7.9  
20:00 234.8 F 9.2  
22:00 241.4 F 11.3  
24:00 242.1 F 11.3  
26:00 237.9 F 9.2  
28:00 251.2 F 15.6  
30:00 262.3 F 21.9  
32:00 272.4 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 27.9  
1:00 272.6 F 28.7  
2:00 271.8 F 27.9  
3:00 272.2 F 28.6  
4:00 272.4 F 28.4  
  
Min 271.4 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.3 F 0.0  
5:00 185.9 F 0.0  
10:00 223.3 F 0.0  
15:00 231.4 F 0.0  
20:00 232.9 F 0.0  
25:00 227.7 F 0.0  
30:00 215.1 F 0.0  
  
Drying Complete

FILLING: 0:42  
HEATING: 32:30  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:10:36  
CYCLE COMPLETE

10/24/2024 09:30 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_