
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 002522
7:13 26/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:06
STERILIZING 3:07
136.5 C 223kPa 3:07
136.4 C 224kPa 3:37
136.4 C 221kPa 4:07
136.4 C 222kPa 4:37
136.4 C 223kPa 5:07
136.2 C 220kPa 5:37
136.3 C 221kPa 6:07
136.4 C 222kPa 6:37
136.4 C 222kPa 6:37
VENTING 6:37
AIR DRYING 7:26
CYCLE COMPLETE 7:29

Digital Signature #
96BD843910232A49
