

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1518

Name: RITTER 1  
Sterilizer ID: V2178198  
01/12/2022 03:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.8 F 0.0  
2:00 169.6 F 3.5  
4:00 187.5 F 7.8  
6:00 214.4 F 13.5  
8:00 223.3 F 7.4  
10:00 236.5 F 12.6  
12:00 236.9 F 9.9  
14:00 236.5 F 8.6  
16:00 249.0 F 14.8  
18:00 261.1 F 21.7  
20:00 272.1 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.9 F 29.2  
2:00 271.5 F 28.4  
3:00 272.9 F 29.4  
4:00 272.3 F 28.9  
  
Min 271.4 F 28.4  
Max 273.1 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.9 F 0.0  
5:00 235.1 F 0.0  
10:00 225.8 F 0.0  
15:00 226.9 F 0.0  
20:00 227.6 F 0.0  
25:00 229.3 F 0.0  
30:00 232.8 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 20:31  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:15  
CYCLE COMPLETE

01/12/2022 04:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_