
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000271
16:22 22/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:08
STERILIZING 3:59
136.0 C 220kPa 3:59
136.3 C 222kPa 4:29
136.4 C 224kPa 4:59
136.2 C 221kPa 5:29
136.4 C 224kPa 5:59
136.1 C 221kPa 6:29
136.4 C 224kPa 6:59
136.2 C 223kPa 7:29
136.2 C 223kPa 7:29
VENTING 7:29
AIR DRYING 8:17
CYCLE COMPLETE 8:50

Digital Signature #
E9D1AAAB2842621F
