
Midmark M11 - v1.0.5
Total Cycles: 3759

Name: RITTER2
Sterilizer ID: V2313228
07/08/2024 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.5 F 0.0
2:00 144.2 F 1.6
4:00 148.3 F 3.9
6:00 181.5 F 7.6
8:00 199.4 F 12.3
10:00 223.7 F 6.0
12:00 231.0 F 10.5
14:00 231.7 F 7.6
16:00 242.4 F 12.3
18:00 239.5 F 10.5
20:00 251.8 F 16.7
22:00 263.1 F 23.1
24:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.4 F 29.4
2:00 272.9 F 29.8
3:00 271.6 F 29.6
4:00 272.0 F 29.3

Min 271.2 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 229.8 F 0.0
10:00 189.4 F 0.0
15:00 187.0 F 0.0
20:00 184.1 F 0.0
25:00 183.0 F 0.0
30:00 182.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:03:27
CYCLE COMPLETE

07/08/2024 12:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____