
Midmark M11 - v1.0.5
Total Cycles: 3624

Name: RITTER2
Sterilizer ID: V2313228
05/16/2024 08:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 79.1 F 0.0
4:00 116.8 F 0.0
6:00 110.4 F 0.2
8:00 147.8 F 1.2
10:00 175.3 F 4.2
12:00 199.9 F 8.1
14:00 213.7 F 12.0
16:00 218.9 F 5.1
18:00 230.3 F 8.9
20:00 239.9 F 13.1
22:00 236.7 F 9.8
24:00 238.1 F 9.4
26:00 246.9 F 14.0
28:00 258.6 F 20.3
30:00 269.1 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.3
2:00 271.7 F 28.9
3:00 272.7 F 29.3
4:00 271.6 F 28.8

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 217.3 F 0.0
10:00 203.3 F 0.0
15:00 191.4 F 0.0
20:00 188.7 F 0.0
25:00 185.7 F 0.0
30:00 198.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:07
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:09
CYCLE COMPLETE

05/16/2024 09:18 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____