
Midmark M11 - v1.0.5
Total Cycles: 3045

Name: RITTER2
Sterilizer ID: V2313228
10/23/2023 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.9 F 0.0
2:00 149.7 F 2.1
4:00 164.8 F 5.5
6:00 190.4 F 10.0
8:00 211.1 F 5.7
10:00 223.2 F 9.1
12:00 232.2 F 9.9
14:00 240.9 F 12.3
16:00 240.8 F 11.3
18:00 254.2 F 18.1
20:00 266.1 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.2 F 29.8
2:00 271.8 F 29.2
3:00 271.5 F 29.5
4:00 272.8 F 30.7

Min 271.0 F 28.9
Max 272.8 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 217.7 F 0.0
10:00 219.6 F 0.0
15:00 216.2 F 0.0
20:00 210.6 F 0.0
25:00 200.5 F 0.0
30:00 197.0 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

10/23/2023 04:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____