
Midmark M11 - v1.0.0
Total Cycles: 3429

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2023 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.5 F 0.0
2:00 205.6 F 3.1
4:00 203.8 F 7.6
6:00 221.0 F 13.0
8:00 227.6 F 8.3
10:00 240.1 F 12.6
12:00 241.0 F 11.4
14:00 239.6 F 10.0
16:00 253.2 F 16.8
18:00 264.7 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.5 F 29.0
2:00 272.1 F 28.5
3:00 271.4 F 28.0
4:00 271.6 F 28.5

Min 271.2 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.9 F 0.0
5:00 233.1 F 0.0
10:00 233.7 F 0.0
15:00 232.3 F 0.0
20:00 215.6 F 0.0
25:00 211.3 F 0.0
30:00 209.0 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

06/22/2023 11:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____