

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001590  
17:24 13/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 1:59  
STERILIZING 3:10  
136.4 C 226kPa 3:10  
136.4 C 226kPa 3:40  
136.3 C 225kPa 4:10  
136.4 C 226kPa 4:40  
136.4 C 226kPa 5:10  
136.4 C 227kPa 5:40  
136.4 C 226kPa 6:10  
136.2 C 224kPa 6:40  
136.2 C 224kPa 6:40  
VENTING 6:40  
AIR DRYING 7:33  
CYCLE COMPLETE 7:46  
-----

Digital Signature #  
F3E9E628ED479E5B  
-----