

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000473  
11:01 12/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:30  
PRESSURIZING 3:31  
STERILIZING 4:41  
136.4 C 233kPa 4:41  
136.1 C 230kPa 5:11  
136.4 C 233kPa 5:41  
136.4 C 233kPa 6:11  
136.3 C 233kPa 6:41  
136.2 C 232kPa 7:11  
136.2 C 232kPa 7:41  
136.4 C 233kPa 8:11  
136.4 C 233kPa 8:11  
VENTING 8:12  
AIR DRYING 9:06  
CYCLE COMPLETE 9:23  
-----

Digital Signature #  
41BA9A5D0A3C173E  
-----