
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002062
8:16 20/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:06
STERILIZING 3:10
136.5 C 223kPa 3:10
136.4 C 224kPa 3:40
136.4 C 224kPa 4:10
136.0 C 222kPa 4:40
136.4 C 225kPa 5:10
136.3 C 224kPa 5:40
136.0 C 221kPa 6:10
136.1 C 223kPa 6:40
136.1 C 223kPa 6:40
VENTING 6:40
AIR DRYING 7:31
CYCLE COMPLETE 13:01

Digital Signature #
388958347E0FC8EC
