
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001756
10:00 01/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:04
STERILIZING 4:17
136.0 C 221kPa 4:17
136.0 C 221kPa 4:47
136.5 C 223kPa 5:17
136.4 C 223kPa 5:47
136.3 C 224kPa 6:17
136.4 C 225kPa 6:47
136.2 C 222kPa 7:17
136.4 C 225kPa 7:47
136.4 C 225kPa 7:47
VENTING 7:47
AIR DRYING 8:38
CYCLE COMPLETE 9:00

Digital Signature #
55832D636361971C
