

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001741  
11:56 17/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:02  
PRESSURIZING 2:07  
STERILIZING 3:10  
136.5 C 221kPa 3:10  
136.4 C 226kPa 3:40  
136.2 C 222kPa 4:10  
136.4 C 225kPa 4:40  
136.1 C 223kPa 5:10  
136.4 C 225kPa 5:40  
136.3 C 223kPa 6:10  
136.2 C 222kPa 6:40  
136.2 C 222kPa 6:40  
VENTING 6:40  
AIR DRYING 7:31  
CYCLE COMPLETE 18:41  
-----

Digital Signature #  
0D3D3BE2FC5C488C  
-----