

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4959

Name: RITTER 1  
Sterilizer ID: V2178198  
10/03/2024 11:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.7 F 0.0  
2:00 157.4 F 0.6  
4:00 177.8 F 2.2  
6:00 217.3 F 6.1  
8:00 212.1 F 10.3  
10:00 230.7 F 7.5  
12:00 231.0 F 8.1  
14:00 240.4 F 12.6  
16:00 238.4 F 9.7  
18:00 239.2 F 9.5  
20:00 247.3 F 13.7  
22:00 258.3 F 19.6  
24:00 268.3 F 25.8  
26:00 272.3 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.7 F 28.9  
2:00 271.8 F 28.3  
3:00 272.4 F 29.0  
4:00 272.4 F 28.6

Min 271.5 F 28.2  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
5:00 211.0 F 0.0  
10:00 244.5 F 0.0  
15:00 231.8 F 0.0  
20:00 208.1 F 0.0  
25:00 185.2 F 0.0  
30:00 176.8 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 26:10  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:04:54  
CYCLE COMPLETE

10/03/2024 12:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_