

-----  
Midmark M11 - v1.0.0  
Total Cycles: 752

Name: RITTER 1  
Sterilizer ID: V2178198  
03/29/2021 06:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.0 F 0.0  
2:00 72.3 F 0.0  
4:00 75.8 F 0.0  
6:00 99.2 F 0.1  
8:00 140.5 F 0.2  
10:00 180.7 F 0.3  
12:00 200.2 F 3.4  
14:00 209.1 F 6.1  
16:00 218.2 F 9.2  
18:00 227.9 F 12.8  
20:00 232.2 F 8.4  
22:00 242.3 F 13.5  
24:00 240.9 F 11.2  
26:00 238.1 F 9.4  
28:00 251.7 F 15.9  
30:00 262.6 F 22.3  
32:00 272.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.4 F 28.5  
2:00 272.3 F 28.9  
3:00 271.5 F 28.1  
4:00 272.6 F 28.8  
  
Min 271.2 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.4 F 0.0  
5:00 187.1 F 0.0  
10:00 199.3 F 0.0  
15:00 195.0 F 0.0  
20:00 187.3 F 0.0  
25:00 192.2 F 0.0  
30:00 189.5 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 32:20  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 01:10:13  
CYCLE COMPLETE

03/29/2021 08:00 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01