
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001150
10:45 30/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:52
STERILIZING 4:00
136.1 C 220kPa 4:00
136.2 C 223kPa 4:30
136.4 C 224kPa 5:00
136.3 C 223kPa 5:30
136.2 C 222kPa 6:00
136.1 C 221kPa 6:30
136.4 C 225kPa 7:00
136.4 C 224kPa 7:30
136.4 C 224kPa 7:30
VENTING 7:31
AIR DRYING 8:19
CYCLE COMPLETE 21:05

Digital Signature #
5089CC81F3F4B71F
