
Midmark M11 - v1.0.5
Total Cycles: 3497

Name: RITTER2
Sterilizer ID: V2313228
04/04/2024 10:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.0 F 0.0
2:00 127.1 F 0.1
4:00 137.9 F 0.7
6:00 174.8 F 4.4
8:00 198.4 F 8.8
10:00 219.8 F 11.4
12:00 222.5 F 8.4
14:00 235.4 F 13.3
16:00 236.8 F 10.6
18:00 235.4 F 8.7
20:00 248.9 F 15.4
22:00 260.7 F 22.0
24:00 271.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.7 F 29.9
2:00 271.6 F 29.1
3:00 272.7 F 30.1
4:00 272.3 F 29.5

Min 271.4 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 227.9 F 0.0
10:00 212.2 F 0.0
15:00 203.5 F 0.0
20:00 198.6 F 0.0
25:00 195.1 F 0.0
30:00 193.6 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:38
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:02:57
CYCLE COMPLETE

04/04/2024 11:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____