
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001880
7:28 23/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:19
PRESSURIZING 2:28
STERILIZING 3:34
136.0 C 220kPa 3:34
136.2 C 223kPa 4:04
136.4 C 224kPa 4:34
136.3 C 222kPa 5:04
136.0 C 222kPa 5:34
136.4 C 224kPa 6:04
136.4 C 224kPa 6:34
136.2 C 223kPa 7:04
136.2 C 223kPa 7:04
VENTING 7:04
AIR DRYING 7:53
CYCLE COMPLETE 37:53

Digital Signature #
EE5BF9933C29BC33
