
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000068
8:20 31/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:17
STERILIZING 3:28
136.0 C 220kPa 3:28
136.1 C 221kPa 3:58
136.2 C 223kPa 4:28
136.3 C 222kPa 4:58
136.1 C 222kPa 5:28
136.4 C 220kPa 5:58
136.2 C 223kPa 6:28
136.2 C 221kPa 6:58
136.2 C 221kPa 6:58
VENTING 6:58
AIR DRYING 7:47
CYCLE COMPLETE 29:07

Digital Signature #
9943402E865E813C
