

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4395

Name: RITTER 1  
Sterilizer ID: V2178198  
03/22/2024 08:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.3 F 0.0  
2:00 70.5 F 0.1  
4:00 77.2 F 1.1  
6:00 107.6 F 3.1  
8:00 133.5 F 5.6  
10:00 163.6 F 8.3  
12:00 190.3 F 10.7  
14:00 195.7 F 5.9  
16:00 214.2 F 5.7  
18:00 225.2 F 9.6  
20:00 234.2 F 10.4  
22:00 238.4 F 9.6  
24:00 238.3 F 9.4  
26:00 245.8 F 12.9  
28:00 257.7 F 19.1  
30:00 267.8 F 25.5  
32:00 272.3 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.5 F 28.7  
2:00 272.1 F 28.7  
3:00 272.2 F 28.5  
4:00 272.3 F 28.9  
  
Min 271.4 F 28.2  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.7 F 0.0  
5:00 227.1 F 0.0  
10:00 197.6 F 0.0  
15:00 201.4 F 0.0  
20:00 193.2 F 0.0  
25:00 207.6 F 0.0  
30:00 190.7 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 32:10  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:10:24  
CYCLE COMPLETE

03/22/2024 09:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_