
Midmark M11 - v1.0.5
Total Cycles: 4190

Name: RITTER2
Sterilizer ID: V2313228
12/05/2024 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.8 F 0.0
2:00 150.1 F 0.3
4:00 167.4 F 2.2
6:00 203.9 F 6.2
8:00 206.3 F 10.7
10:00 222.8 F 6.4
12:00 230.2 F 10.4
14:00 232.8 F 8.5
16:00 242.6 F 13.3
18:00 242.7 F 12.4
20:00 254.7 F 18.6
22:00 265.5 F 25.1
24:00 271.0 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.8
1:00 272.0 F 29.7
2:00 272.2 F 29.7
3:00 272.7 F 29.9
4:00 272.2 F 30.1

Min 271.7 F 29.7
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.6 F 0.0
5:00 211.9 F 0.0
10:00 190.9 F 0.0
15:00 190.2 F 0.0
20:00 188.4 F 0.0
25:00 185.9 F 0.0
30:00 183.5 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:33
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:03:31
CYCLE COMPLETE

12/05/2024 02:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____