

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005784  
13:20 19/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:25  
PRESSURIZING 2:27  
STERILIZING 3:32  
136.4 C 223kPa 3:32  
136.4 C 225kPa 4:02  
136.4 C 225kPa 4:32  
136.1 C 220kPa 5:02  
136.4 C 224kPa 5:32  
136.2 C 222kPa 6:02  
136.4 C 225kPa 6:32  
136.2 C 221kPa 7:02  
136.2 C 221kPa 7:02  
VENTING 7:02  
AIR DRYING 7:58  
CYCLE COMPLETE 8:02

-----  
Digital Signature #  
7079AE5DB01F5637  
-----