

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1166

Name: RITTER 1  
Sterilizer ID: V2178198  
09/14/2021 06:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	73.4 F	0.0
2:00	78.2 F	0.0
4:00	95.3 F	0.0
6:00	130.1 F	0.0
8:00	163.9 F	0.1
10:00	178.6 F	0.2
12:00	195.0 F	0.3
14:00	207.2 F	3.1
16:00	216.2 F	5.6
18:00	225.5 F	8.5
20:00	234.4 F	12.0
22:00	231.2 F	7.6
24:00	242.0 F	12.6
26:00	239.9 F	10.5
28:00	236.8 F	8.7
30:00	250.3 F	15.1
32:00	261.1 F	21.3
34:00	271.0 F	27.7

STERILIZING

mm:ss	Degrees	PSI
0:00	271.9 F	28.0
1:00	271.7 F	28.3
2:00	272.7 F	28.8
3:00	272.2 F	28.3
4:00	271.4 F	28.0

Min 271.4 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	177.0 F	0.0
	5:00	191.0 F	0.0
	10:00	190.6 F	0.0
	15:00	191.4 F	0.0
	20:00	201.5 F	0.0
Drying Complete	25:00	184.4 F	0.0
	30:00	190.0 F	0.0

FILLING: 0:52  
HEATING: 34:39  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:12:52  
CYCLE COMPLETE

09/14/2021 07:48 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01