
Midmark M11 - v1.0.5
Total Cycles: 3307

Name: RITTER2
Sterilizer ID: V2313228
01/19/2024 09:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 81.7 F 0.0
4:00 134.0 F 0.0
6:00 103.2 F 1.4
8:00 140.7 F 4.5
10:00 169.2 F 8.1
12:00 205.0 F 12.6
14:00 216.9 F 5.3
16:00 229.4 F 9.2
18:00 240.6 F 12.8
20:00 236.0 F 9.8
22:00 237.2 F 9.0
24:00 246.4 F 13.8
26:00 258.6 F 20.4
28:00 269.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.3 F 29.3
2:00 271.8 F 29.3
3:00 272.1 F 29.0
4:00 272.3 F 29.7

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 237.0 F 0.0
10:00 195.5 F 0.0
15:00 187.1 F 0.0
20:00 184.1 F 0.0
25:00 184.5 F 0.0
30:00 195.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:02
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:06:49
CYCLE COMPLETE

01/19/2024 10:16 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____