

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2919

Name: RITTER2  
Sterilizer ID: V2313228  
09/25/2023 12:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.2 F 0.0  
2:00 159.0 F 1.2  
4:00 174.1 F 4.6  
6:00 193.8 F 8.7  
8:00 213.0 F 13.1  
10:00 225.1 F 7.5  
12:00 239.0 F 12.2  
14:00 233.6 F 9.3  
16:00 238.3 F 10.4  
18:00 245.6 F 13.6  
20:00 256.7 F 19.6  
22:00 267.2 F 25.9  
24:00 272.4 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.3 F 29.3  
2:00 272.1 F 29.7  
3:00 271.5 F 29.7  
4:00 271.4 F 29.2

Min 271.1 F 29.1  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.2 F 0.0  
5:00 192.3 F 0.0  
10:00 191.4 F 0.0  
15:00 191.1 F 0.0  
20:00 189.1 F 0.0  
25:00 188.2 F 0.0  
30:00 200.2 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 24:10  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:02:60  
CYCLE COMPLETE

09/25/2023 01:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_