

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2238

Name: RITTER 1  
Sterilizer ID: V2178198  
08/24/2022 10:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.0 F 0.0  
2:00 172.2 F 0.2  
4:00 179.6 F 0.9  
6:00 202.8 F 4.2  
8:00 222.2 F 8.6  
10:00 233.5 F 13.5  
12:00 235.4 F 9.9  
14:00 236.3 F 8.8  
16:00 246.3 F 13.5  
18:00 244.6 F 12.3  
20:00 257.0 F 18.9  
22:00 267.9 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.4 F 28.7  
2:00 272.0 F 28.6  
3:00 272.0 F 28.4  
4:00 272.3 F 28.9

Min 271.4 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 219.5 F 0.0  
10:00 235.0 F 0.0  
15:00 226.0 F 0.0  
20:00 223.9 F 0.0  
25:00 221.9 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 23:21  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:01:24  
CYCLE COMPLETE

08/24/2022 11:22 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_