
Midmark M11 - v1.0.0
Total Cycles: 1056

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2021 06:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 78.2 F 0.0
4:00 83.9 F 0.0
6:00 121.6 F 0.1
8:00 155.3 F 0.2
10:00 184.0 F 0.3
12:00 201.9 F 3.0
14:00 214.0 F 6.0
16:00 223.3 F 9.2
18:00 233.1 F 12.9
20:00 233.0 F 8.5
22:00 244.1 F 12.9
24:00 242.3 F 11.5
26:00 239.4 F 10.0
28:00 253.0 F 16.7
30:00 264.3 F 23.3
32:00 271.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.8
2:00 271.5 F 28.1
3:00 272.8 F 28.9
4:00 271.7 F 28.0

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 197.8 F 0.0
10:00 178.7 F 0.0
15:00 176.9 F 0.0
20:00 186.5 F 0.0
25:00 188.0 F 0.0
30:00 187.5 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:00
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:58
CYCLE COMPLETE

07/27/2021 07:42 AM
ACCEPTED: X
REJECTED:
Operator: U01