
Midmark M11 - v1.0.5
Total Cycles: 266

Name: RITTER2
Sterilizer ID: V2313228
08/05/2021 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.4 F 0.0
2:00 165.5 F 2.8
4:00 178.2 F 6.7
6:00 195.0 F 11.6
8:00 218.1 F 5.6
10:00 229.8 F 10.6
12:00 231.3 F 7.6
14:00 244.1 F 13.3
16:00 243.0 F 12.2
18:00 255.3 F 18.5
20:00 266.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.6 F 29.7
2:00 272.3 F 29.0
3:00 271.4 F 28.9
4:00 272.4 F 29.7

Min 271.2 F 28.7
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.0 F 0.0
5:00 229.3 F 0.0
10:00 222.0 F 0.0
15:00 228.6 F 0.0
20:00 227.8 F 0.0
25:00 221.6 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 21:44
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

08/05/2021 11:33 AM
ACCEPTED: X
REJECTED:
Operator: U02