

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1923

Name: RITTER 1  
Sterilizer ID: V2178198  
05/10/2022 09:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.3 F 0.0  
2:00 149.9 F 0.6  
4:00 160.2 F 2.7  
6:00 183.4 F 5.9  
8:00 209.5 F 10.4  
10:00 217.3 F 4.5  
12:00 230.1 F 9.3  
14:00 238.6 F 10.1  
16:00 240.5 F 11.2  
18:00 237.2 F 8.9  
20:00 250.8 F 15.4  
22:00 262.5 F 22.1  
24:00 272.7 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.6 F 28.8  
2:00 272.0 F 28.2  
3:00 271.5 F 28.2  
4:00 272.7 F 28.8

Min 271.4 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 231.4 F 0.0  
10:00 230.9 F 0.0  
15:00 231.3 F 0.0  
20:00 227.2 F 0.0  
25:00 232.3 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 24:21  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:02:14  
CYCLE COMPLETE

05/10/2022 10:15 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_