

-----  
Midmark M11 - v1.0.5  
Total Cycles: 922

Name: RITTER2  
Sterilizer ID: V2313228  
03/29/2022 01:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 103.3 F 0.0  
2:00 107.2 F 0.0  
4:00 127.3 F 0.1  
6:00 165.7 F 0.4  
8:00 191.5 F 3.9  
10:00 223.3 F 8.1  
12:00 232.2 F 12.9  
14:00 235.3 F 9.3  
16:00 236.8 F 10.0  
18:00 244.6 F 13.2  
20:00 243.8 F 12.4  
22:00 256.8 F 19.3  
24:00 268.0 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.4 F 29.6  
2:00 271.4 F 28.7  
3:00 272.1 F 29.7  
4:00 272.2 F 29.3

Min 271.1 F 28.7  
Max 272.4 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.5 F 0.0  
5:00 201.5 F 0.0  
10:00 209.6 F 0.0  
15:00 214.7 F 0.0  
20:00 211.1 F 0.0  
25:00 209.1 F 0.0  
30:00 212.1 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:17  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:03:20  
CYCLE COMPLETE

03/29/2022 02:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_