
Midmark M11 - v1.0.0
Total Cycles: 5049

Name: RITTER 1
Sterilizer ID: V2178198
11/04/2024 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.9 F 0.0
2:00 159.1 F 0.2
4:00 166.0 F 0.9
6:00 188.2 F 4.9
8:00 206.3 F 9.1
10:00 220.7 F 13.4
12:00 226.8 F 8.3
14:00 239.5 F 13.1
16:00 239.9 F 10.2
18:00 236.7 F 8.4
20:00 249.4 F 14.6
22:00 260.4 F 20.8
24:00 270.5 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.5 F 28.6
2:00 272.7 F 29.0
3:00 271.6 F 28.6
4:00 272.5 F 28.7

Min 271.4 F 28.4
Max 273.1 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 230.3 F 0.0
10:00 251.8 F 0.0
15:00 238.9 F 0.0
20:00 230.5 F 0.0
25:00 229.6 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 25:28
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:03
CYCLE COMPLETE

11/04/2024 12:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____