
Midmark M11 - v1.0.0
Total Cycles: 4850

Name: RITTER 1
Sterilizer ID: V2178198
08/26/2024 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 68.5 F 0.0
4:00 80.5 F 0.0
6:00 116.4 F 0.2
8:00 147.3 F 0.3
10:00 171.3 F 0.5
12:00 191.7 F 0.8
14:00 204.8 F 1.6
16:00 218.9 F 4.8
18:00 224.6 F 7.8
20:00 233.7 F 11.2
22:00 232.1 F 7.0
24:00 242.8 F 12.0
26:00 238.7 F 9.7
28:00 238.4 F 9.0
30:00 248.8 F 14.3
32:00 259.5 F 20.2
34:00 269.1 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 27.9
1:00 271.7 F 28.0
2:00 272.7 F 28.7
3:00 272.6 F 28.5
4:00 272.0 F 28.0

Min 271.5 F 27.8
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 188.7 F 0.0
10:00 186.7 F 0.0
15:00 192.4 F 0.0
20:00 217.3 F 0.0
25:00 234.5 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:41
HEATING: 35:07
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:13:18
CYCLE COMPLETE

08/26/2024 10:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____