
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002989
15:08 19/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:39
STERILIZING 3:38
136.4 C 225kPa 3:38
136.4 C 224kPa 4:08
136.4 C 225kPa 4:38
136.4 C 223kPa 5:08
136.4 C 223kPa 5:38
136.4 C 222kPa 6:08
136.4 C 225kPa 6:38
135.9 C 219kPa 7:08
135.9 C 219kPa 7:08
VENTING 7:08
AIR DRYING 7:57
CYCLE COMPLETE 8:01

Digital Signature #
927A60EA4E598BE6
