
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001045
8:54 28/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 4:12
STERILIZING 5:44
136.0 C 221kPa 5:44
136.3 C 224kPa 6:15
136.4 C 224kPa 6:45
136.4 C 224kPa 7:15
136.4 C 225kPa 7:45
136.2 C 220kPa 8:15
136.3 C 224kPa 8:45
136.4 C 225kPa 9:15
136.4 C 225kPa 9:15
VENTING 9:15
AIR DRYING 10:11
CYCLE COMPLETE 25:11

Digital Signature #
513B2A25C39C0E9C
