
Midmark M11 - v1.0.5
Total Cycles: 1480

Name: RITTER2
Sterilizer ID: V2313228
09/13/2022 06:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.0 F 0.0
2:00 97.2 F 0.0
4:00 147.6 F 0.0
6:00 170.0 F 0.1
8:00 202.3 F 0.3
10:00 214.4 F 3.7
12:00 228.4 F 7.7
14:00 229.8 F 11.9
16:00 227.8 F 5.7
18:00 237.5 F 10.0
20:00 239.5 F 10.1
22:00 242.5 F 11.8
24:00 239.1 F 10.1
26:00 251.8 F 16.4
28:00 262.6 F 22.7
30:00 272.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.7 F 29.3
2:00 272.7 F 29.3
3:00 272.7 F 29.4
4:00 272.8 F 29.6

Min 271.0 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.0 F 0.0
5:00 240.2 F 0.0
10:00 212.1 F 0.0
15:00 213.3 F 0.0
20:00 215.8 F 0.0
25:00 221.0 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 30:24
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:08:30
CYCLE COMPLETE

09/13/2022 07:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____