
Midmark M11 - v1.0.0
Total Cycles: 3070

Name: RITTER 1
Sterilizer ID: V2178198
04/03/2023 03:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.3 F 0.0
2:00 157.5 F 1.6
4:00 185.4 F 4.8
6:00 209.0 F 8.9
8:00 216.1 F 8.5
10:00 226.1 F 8.4
12:00 238.7 F 12.5
14:00 237.5 F 10.7
16:00 236.0 F 8.7
18:00 250.0 F 15.6
20:00 262.1 F 22.5
22:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.6 F 29.4
2:00 271.7 F 28.6
3:00 272.1 F 29.3
4:00 272.3 F 29.2

Min 271.2 F 28.5
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 208.9 F 0.0
10:00 232.7 F 0.0
15:00 241.2 F 0.0
20:00 227.6 F 0.0
25:00 222.5 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 22:21
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

04/03/2023 04:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____