

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1013

Name: RITTER 1  
Sterilizer ID: V2178198  
07/14/2021 02:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.5 F 0.0  
2:00 190.4 F 3.6  
4:00 203.5 F 8.5  
6:00 222.9 F 12.6  
8:00 229.9 F 9.2  
10:00 238.1 F 9.7  
12:00 242.9 F 12.6  
14:00 241.9 F 11.4  
16:00 255.0 F 18.2  
18:00 266.5 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.5 F 28.9  
2:00 272.4 F 29.1  
3:00 272.4 F 28.7  
4:00 272.4 F 29.2  
  
Min 271.6 F 28.4  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.5 F 0.0  
5:00 234.3 F 0.0  
10:00 233.1 F 0.0  
15:00 227.2 F 0.0  
20:00 224.2 F 0.0  
25:00 221.3 F 0.0  
30:00 218.8 F 0.0  
Drying Complete

FILLING: 2:24  
HEATING: 19:31  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:06  
CYCLE COMPLETE

07/14/2021 03:21 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01