

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000279  
14:43 01/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:43  
PRESSURIZING 3:44  
STERILIZING 5:04  
136.0 C 229kPa 5:04  
136.3 C 233kPa 5:34  
136.4 C 233kPa 6:04  
136.4 C 233kPa 6:34  
136.4 C 233kPa 7:04  
136.4 C 233kPa 7:34  
136.4 C 232kPa 8:04  
136.4 C 232kPa 8:34  
136.4 C 232kPa 8:34  
VENTING 8:34  
AIR DRYING 9:40  
CYCLE COMPLETE 22:40  
-----

Digital Signature #  
80DC05956D649892  
-----